

A novel stereoscopic projection display system for CT images of fractures

XIUJUAN LIU¹, HONG JIANG², YUEDONG LANG³, HONGBO WANG⁴ and NA SUN¹

¹Department of CT Room, 1st Affiliated Hospital of Harbin Medical University, Harbin;

²Department of CT Room, Heilongjiang Province Hospital Nangang Branch, Harbin, Heilongjiang 150001;

³Beijing Aerospace Control Instrument Research Institute, Beijing 100854;

⁴School of Mechatronics, Harbin Institute of Technology, Harbin, Heilongjiang 150001, P.R. China

Received January 11, 2013; Accepted March 22, 2013

DOI: 10.3892/etm.2013.1044

Abstract. The present study proposed a novel projection display system based on a virtual reality enhancement environment. The proposed system displays stereoscopic images of fractures and enhances the computed tomography (CT) images. The diagnosis and treatment of fractures primarily depend on the post-processing of CT images. However, two-dimensional (2D) images do not show overlapping structures in fractures since they are displayed without visual depth and these structures are too small to be simultaneously observed by a group of clinicians. Stereoscopic displays may solve this problem and allow clinicians to obtain more information from CT images. Hardware with which to generate stereoscopic images was designed. This system utilized the conventional equipment found in meeting rooms. The off-axis algorithm was adopted to convert the CT images into stereo image pairs, which were used as the input for a stereo generator. The final stereoscopic images were displayed using a projection system. Several CT fracture images were imported into the system for comparison with traditional 2D CT images. The results showed that the proposed system aids clinicians in group discussions by producing large stereoscopic images. The results demonstrated that the enhanced stereoscopic CT images generated by the system appear clearer and smoother, such that the sizes, displacement and shapes of bone fragments are easier to assess. Certain fractures that were previously not visible on 2D CT images due to vision overlap became vividly evident in the stereo images. The proposed projection display system efficiently, economically and accurately displayed three-dimensional (3D) CT images. The system may help clinicians improve the diagnosis and treatment of fractures.

Introduction

The stereoscopic display of fractures is required by clinicians since the diagnosis and treatment of fractures are highly dependent on computed tomography (CT) post-processing images for manual reduction. Shaded volume rendering (SVR) is commonly used in hospitals as an effective and convenient method for reconstructing stereo CT images in CT workstations. Three-dimensional (3D) models may be reconstructed from sequential slices via SVR and shown on two-dimensional (2D) screens or medical films; the structures are made translucent to allow the clinicians to observe the interior of the bone (1,2). However, these 3D images are too small to be simultaneously observed by a group of clinicians. In addition, the visual data regarding the mutual location, size and shape of bone fragments in fractures often overlap or are not visible since these images are displayed without visual depth. Therefore, the evaluation of complex anatomical regions requires considerable experience from clinicians and involves several potential difficulties (3).

The 3D structures and shapes from volumetric data are difficult to perceive in translucent volumes without depth cues (4), whereas a feeling of depth during monitoring may provide more information (5). To overcome these difficulties, Kniss *et al* (6) combined volume rendering with a virtual reality display. Shen *et al* (7) described a real-time medical visualization system for analyzing volumetric data in various virtual environments. In the system provided by Ji *et al* (8), stereo viewing was achieved using a head-mounted display (HMD). Zhang *et al* (9) proposed a method for stereo image generation from video data using phase correlation technology. Without 3D information, respective motion parameters for the formation of a stereo pair for the left and right eyes were computed, from which the stereo images were generated (8). Zinger *et al* (10) presented another approach for the autostereoscopic display of medical images and developed an efficient algorithm for view interpolation and rendering, which was based on texture-depth data representation. In addition to the virtual reality methods, mixed-reality methods have been applied for medical image displays. Ferrari *et al* (11) generated augmented images to provide users with a stereoscopic view. Sielhorst (12) studied the medical application of depth

Correspondence to: Professor Na Sun, Department of CT Room, 1st Affiliated Hospital of Harbin Medical University, 199 Dongdazhi Street, Nangang, Harbin, Heilongjiang 150001, P.R. China
E-mail: sunnacn@126.com

Key words: projection, stereo display, computed tomography image, fracture

perception using translucent head-mounted displays (HMDs). These studies were conducted in laboratories, but medical applications of a technology have more practical requirements. To provide more accessible CT or MRI medical images to patients, Okuyama *et al.* (13) reconstructed stereoscopic CT and MRI images and displayed them on a notebook personal computer by adopting binocular-type stereoscopy. These studies enabled simple and low-cost autostereoscopy with a small-scale system (14). However, this desktop system does not fulfill the requirements of cooperative work among doctors during diagnosis.

The present study presents a novel method of projection display based on a virtual reality enhancement environment system (VREES). The proposed system stereoscopically displays fractures with the advantages of clear images and low operating costs. This system converted the post-processing 2D images from CT workstations into large 3D images, which may be simultaneously displayed to several clinicians for discussion and consultation. Therefore, clinicians are able to obtain more information from the 3D images than from the 2D images and treat patients with fractures with greater efficacy.

Materials and methods

Hardware and procedure. Active stereo projection was adopted as the display method due to its generality, feasibility and cost. A stereo generating device was designed to replace the advanced graphics cards and to decrease the system costs. In this system, the image resources were obtained by the off-axis algorithm and synthesized as a time-sequenced image pair by a 3D image generator. The image pair were projected onto the screen and the control signals were transmitted to a 3D control valve to drive the shutter of the 3D glasses. The modules of the system are shown in Fig. 1.

One advantage of the proposed system is that only low-level drivers are required for stereo display, instead of professional graphics cards. The key problem was the creation of the image pair for the left and right eyes in real-time. The medical images were automatically saved by the CT workstation as a sequence with different viewpoints that may be used as input images. The off-axis algorithm transformed the source images into stereo pairs in the software module of the proposed system. Therefore, stereo projection display was achieved with a standard personal computer and a consumer-level stereo projector.

The following equipment was used in the VREES to cost-effectively and efficiently display 3D images; a computer, a large display screen, a projector, a 3D-image generator, 3D glasses and 3D control emitters. The equipment may be purchased at reasonable prices, with the exception of the 3D image generator, which was manufactured.

The 3D image generator in our proposed system was composed of five parts: an input-and-output interface, a power interface, a pulse-generating unit, an image-processing unit and a 3D-driving unit. Three tasks had to be completed by the system while the generator was operational. The first task was to obtain the ordered images according to the pulse signals from the pulse generating unit and to feed these images into the 3D-image generator according to the same time sequence. The second task was to generate 3D signals according to the time rules of the vertical sync signal from the pulse generating




Figure 1. Three dimensional (3D) projection system based on 3D image generator.




Figure 2. Control circuit principle of three dimensional (3D) image generator.

unit and to alternately translate these signals into the next unit. The third task was to produce the driving 3D control signal, which is used to synchronize the opening and closing of the shutters of the 3D glasses with the vertical sync signal.

The control circuit of the 3D image generator is shown in Fig. 2. Images or videos were used as sources of the 3D signals, which were sent to the generator through the left and right input interfaces. These left and right signals were converted to the RGB format by the input interface. Two signals were then synchronously selected and controlled with a specific time sequence in the pulse generating unit.

The workflow of the generator in the proposed system had five steps. i) The 3D image generator first obtained the input image or video signals and estimated whether the refresh frequencies of the two-way signals were synchronous. If the refresh frequencies were not synchronous, the ongoing procedure was terminated and the graphics card started the refresh synchronization mode. ii) When the refresh frequencies were synchronous, the video synchronization mode opened the switch control of the two-way image signals. iii) The image processing unit then created a 3D image signal for the high-brightness projector. iv) Once the generator finished creating the appropriate 3D image signal, the video graphics array (VGA) interface exported this signal to a high-brightness projector, whereas a special interface exported the 3D synchronous control signal to the infrared emitter of the stereo



Figure 4. Use of a VRS by a single doctor. VRS, virtual reality system.




Figure 5. Application of the VREES for discussion and consultation. VREES, virtual reality enhancement environment system.




Figure 6. Medial orbital walls in the right eye with local defects (circle) and inferior orbital wall fracture (rectangle).




Figure 7. Fracture without displacement.

Results

Large-scale. Compared with conventional virtual reality system (VRS) displays, the 3D visualization of fractures using




Figure 8. Displacement, rotation and distance of fractured ends.

VREES was observed to be a more convenient and efficient method for groups of clinicians to diagnose and discuss fractures. As shown in Fig. 4, a single clinician operated the computer for observing the CT image on the VRS desktop. The resulting 2D images were too small to be simultaneously observed and discussed by several clinicians. By contrast, a maximum of seven clinicians with 3D glasses observed the 3D images clearly on a large scale when using the VREES. Each clinician received identical and complete information from the 3D images (Fig. 5).

3D display of fractures with local defects. To illustrate the advantages of the 3D images of fractures generated by the VREES, several applications of the proposed system on fracture images were performed. Certain parts of the lamina cribrosa in the medial orbital wall of the eye were too thin to be selected as bone volume elements within the limits of the CT values in SVR. Therefore, these parts were not shown in the CT images. The complete medial orbital wall with its local defects was not observed by the ophthalmologists. As shown in Fig. 6 (circle), the images of the medial orbital walls with local defects in the eye overlapped and were not clear. These walls were close to each other. Therefore, the 2D images lacked visual depth. The shape of the orbital walls, the fracture itself, and the placement of the fracture along the medial orbital wall in the eye were difficult for clinicians to observe and diagnose. Consequently, the information remained hidden in the 2D images. By contrast, when the visual depth of the image was reconstructed in the VREES, the two medial orbital walls of the eye were displayed distinctly. The fracture line and the sclerotic with local defects as well as the size and displacement of the bone fragments were clearly visible. With the aid of 3D images in the VREES, ophthalmologists were able to calculate the quantity and shape of the artificial bone pieces. These pieces were used to improve the accuracy of filling in local defects of the sclerotic and the positioning of titanium plates on fractures.

The inferior orbital wall, which is opposite the upper maxillary sinus wall, had an uneven thickness in the obtained images. The CT images showed that the thinness of this wall was similar to that of the medial orbital wall in the eye. In addition, the orbital fossa formed tetrahedron and the 2D images overlapped without visual depth. Thus, the sclerotic of the inferior orbital wall was not clearly visible in the 2D images, as shown in Fig. 6 (rectangle). By contrast, the sclerotic of the inferior orbital wall and the fracture were

clearly displayed with the reconstruction of visual depth in the VREES.

3D display of fractures without displacement. Different types of human tissue were shown in SVR to have varying diaphaneity. When one type of tissue was observed through a different tissue, the 2D images of these two tissues obscured each other. Consequently, fractures without displacement or those with a displacement of <2 mm were generally unclear, as shown in Fig. 7 where the fracture is not visible. However, fracture lines that were not visible in the 2D images were clearly and vividly displayed in the VREES since the 3D images showed the stereo view of the bone to the orthopedists. The orthopedists were able to observe the fracture from various directions, such that all signs of the fracture were detectable. This ability to observe fractures from any direction is of great significance to orthopedists for the diagnosis and evaluation of fractures, including those that are extremely small.

3D display of fractures with displacement, rotation and distance of fracture ends. The displacement of a fracture in the intra-articular, front and back regions, as well as the size of the bone fragments and the fracture site, were all presented in the single, unclear 2D image that was generated in SVR without visual depth, as shown in Fig. 8. To obtain more information on a fracture, doctors were able to rotate the 2D images on-screen with SVR. However, the doctors were then required to mentally reconstruct a 3D model of the fracture site. Therefore, the information obtained by one doctor may differ from that obtained by others, which may make it difficult for the doctors to agree. By contrast, the 3D fracture information was clearly shown in the VREES. The doctors obtained identical information regarding the articular cavity, thereby enabling them to easily come to an agreement and reasonably select the most suitable surgical approach. Furthermore, the doctors were able to effectively use the fixation method for accurate fracture reduction.

Discussion

Stereo CT images of fractures are commonly displayed on a plate screen using the volume element reconstruction method (16). The stereo projection display that is featured in the VREES produced images in a virtual reality system that were able to help clinicians diagnose various conditions, plan treatment options and monitor changes over time as well as predict and display final treatment results (6,17). Applications of the proposed system showed that the visual depth was reconstructed in the 3D images that were produced in the VREES. Stereo CT images were clearly displayed by the proposed system, including the size, shape and displacement of bone fragments as well as the front and back displacements of the fracture. Certain fractures were not visible in the 2D CT images due to image overlap. However, these fractures were vividly displayed in the VREES. These results are consistent with those of previous virtual reality methods (9,10). The proposed system efficiently meets the requirements of clinicians for fracture observation, but also provides a large visual scope for orthopedists to conveniently discuss fractures in groups.

Numerous software systems are installed in CT machines to provide 3D images on flat screens. Clinicians generally

observe 3D models on 2D display devices. If clinicians wish to see planar 3D images in their respective departments, the special software that is required for CT machines is unavailable due to prohibitive pricing.

Stereo projection display systems consist of a computer, a large screen or white wall, a projector, a 3D image generator, 3D glasses and 3D control emitters, the majority of which are conventional devices in meeting rooms (18). HMDs are coupled with 3D perspective-rendering algorithms in mixed-reality systems (19,20). By contrast, stereo projection display systems use fewer and less expensive devices. Thus, stereo projection display systems are relatively economical, thereby allowing clinicians in underdeveloped areas to clearly observe the fractures of their patients as easily as their counterparts in developed areas.

A stereo projection display system is just one method for clinicians to observe 3D CT images. Clinicians usually focus on the accuracy and convenience of the systems being used. However, the source images of the proposed system are derived from CT post-processing images, which are the original CT images without modification. Thus, the information obtained by clinicians from the 3D images in the proposed system is similar to that from the original CT post-processing images. Therefore, the accuracy of the proposed system is undisputed. However, clinicians are often uncomfortable with wearing 3D glasses. If 3D images may be observed without 3D glasses, clinicians may find the system more comfortable and convenient. Therefore, future studies on a system that does not require 3D glasses are required.

Acknowledgements

This study was supported by the Natural Science Foundation from natural science fund committee of Heilongjiang Province, China (D2007-47).

References

1. Wang XH, Good WF, Fuhrman CR, Sumkin JH, Britton CA and Golla SK: Stereo CT image compositing methods for lung nodule detection and characterization. *Acad Radiol* 12: 1512-1520, 2005.
2. Brunner A, Honigsmann P, Treumann T and Babst R: The impact of stereo-visualisation of three dimensional CT datasets on the inter- and intraobserver reliability of the AO/OTA and Neer classifications in the assessment of fractures of the proximal humerus. *J Bone Joint Surg Br* 91: 766-771, 2009.
3. Ringl H, Scherthaner R, Philipp MO, et al: Three-dimensional fracture visualisation of multidetector CT of the skull base in trauma patients: comparison of three reconstruction algorithms. *Eur Radiol* 19: 2416-2424, 2009.
4. Kersten MA, Stewart AJ, Troje N and Ellis R: Enhancing depth perception in translucent volumes. *IEEE Trans Vis Comput Graph* 12: 1117-1123, 2006.
5. Kim JH, Lee JY and Lee MH: Development of 3-D stereo endoscopic image processing system. *Inst Electron Inf Commun Eng E85-D: 584-591*, 2002.
6. Kniss J, Schulze JP, Wössner U, Winkler P, Lang U and Hansen C: Medical applications of multi-field volume rendering and VR techniques. In: *VISSYM'04 Proceedings of the Sixth Joint Eurographics - IEEE TCVG Conference on Visualization*. pp249-254, 2004.
7. Shen R, Boulanger P and Noga M: MedVis: A real-time immersive visualization environment for the exploration of medical volumetric data. In: *Proceedings of the Fifth International Conference on BioMedical Visualization (MediVis 2008): Information Visualization in Medical and Biomedical Informatics*. pp63-68, 2008.

8. Ji W, Williams RL, Howell JN and Conatser RR Jr: 3D stereo viewing evaluation for the virtual haptic back project. In: Proceedings of the 14th International Symposium on Haptic Interfaces for Virtual Environment and Teleoperator Systems (HAPTICS 2006). pp251-258, 2006.
9. Zhang XH and Nakajima M: Visual stereo image generation from video data using phase correlation technique. IEICE Transactions on Information and Systems E83-D: 1266-1273, 2000.
10. Zinger S, Ruijters D, Do L and Peter HN: View interpolation for medical images on autostereoscopic displays. IEEE Transactions on Circuits and Systems for Video Technology 22: 128-137, 2012.
11. Ferrari V, Megali G, Troia E, Pietrabissa A and Mosca F: A 3-D mixed-reality system for stereoscopic visualization of medical dataset. IEEE Trans Biomed Eng 56: 2627-2633, 2009.
12. Sielhorst T, Bichlmeier C, Heining SM and Navab N: Depth perception - a major issue in medical AR: evaluation study by twenty surgeons. Med Image Comput Comput Assist Interv 9: 364-372, 2006.
13. Okuyama F, Sugase T and Hirano T: Application of PC stereoscopic image viewer for informed consent. In: Proceedings of the Fourth International Conference on Computer and Information Technology, 2004. CIT '04. pp902-906, 2004.
14. Vogt F and Wagner AY: Stereo pairs in astrophysics. Astrophysics Space Sci 337: 79-92, 2012.
15. Stenholt R: Stereo Rendering. http://www.cvmt.dk/education/teaching/e07/cvg9/VR/Stereo_Rendering.pdf. Accessed March 26, 2013.”.
16. Mitton D, Zhao K, Bertrand S, *et al*: 3D reconstruction of the ribs from lateral and frontal X-rays in comparison to 3D CT-scan reconstruction. J Biomech 41: 706-710, 2008.
17. Liu Y, Olszewski R, Alexandroni ES, Enciso R, Xu T and Mah JK: The validity of in vivo tooth volume determinations from cone-beam computed tomography. Angle Orthod 80: 160-166, 2010.
18. Patel D, Muren LP, Mehus A, Kvinnsland Y, Ulvang DM and Villanger KP: A virtual reality solution for evaluation of radiotherapy plans. Radiother Oncol 82: 218-221, 2007.
19. Looser J, Billinghamurst M and Cockburn A: Through the looking glass: the use of lenses as an interface tool for augmented reality interfaces. In: Proceedings of the 2nd international Conference on Computer Graphics and interactive Techniques in Australasia and South East Asia (GRAPHITE 2004). ACM, New York, NY, pp204-211, 2004.
20. Kalkofen D, Mendez E and Schmalstieg D: Interactive focus and context visualization for augmented reality. In: Proceedings of the Sixth IEEE and ACM International Symposium on Mixed and Augmented Reality. pp191-201, 2007.