

A novel optical coherence tomography-based calcium scoring system can predict the stent expansion of moderate and severe calcified lesions

CHANG HOU^{1,2*}, LINJIAN YANG^{1,2*}, ZIXUAN XUE^{1,2}, HAIMIAO LIN^{1,2}, YULIANG MA^{1,2},
QI LI^{1,2}, CHUANFEN LIU^{1,2}, MINGYU LU^{1,2}, HONG ZHAO^{1,2} and JIAN LIU^{1,2}

¹Department of Cardiology and ²Beijing Key Laboratory of Early Prediction and Intervention of Acute Myocardial Infarction, Peking University People's Hospital, Beijing 100044, P.R. China

Received April 4, 2022; Accepted September 16, 2022

DOI: 10.3892/etm.2022.11667

Abstract. Coronary calcified lesions can exert serious effects on stent expansion. A calcium scoring system, based on optical coherence tomography (OCT), has been previously developed to identify relatively mild calcified lesions that would benefit from plaque modification procedures. Therefore, the present study aimed to establish a novel OCT-based scoring system to predict the stent expansion of moderate and severe calcified lesions. A total of 33 patients who underwent percutaneous coronary intervention (PCI; 34 calcified lesions were observed using coronary angiography) were retrospectively included in the present study. Coronary angiography and OCT images were subsequently reviewed and analyzed. Furthermore, a calcium scoring system was developed based on the results of multivariate analysis before the optimal threshold for the prediction of stent underexpansion in patients with moderate and severe calcified lesions was determined. The mean age of the patients was 67±10 years. The present analysis demonstrated that the final post-PCI median stent expansion was 70.74%, where stent underexpansion (defined as stent expansion <80%) was observed in 23 lesions. The mean maximum calcium arc, length and thickness, which were assessed using OCT, were found to be 230°, 25.10 mm and 1.18 mm, respectively. A multivariate logistic regression model demonstrated that age and the maximum calcium arc were independent predictors of stent underexpansion. A novel calcium scoring system was thereafter established using the following formula: $(0.16 \times \text{age}) + (0.03 \times \text{maximum calcium arc})$ according to the

β -coefficients in the multivariate analysis, with the optimal cut-off value for the prediction of stent underexpansion being 16.87. Receiver operating characteristic curve analysis demonstrated that this novel scoring system yielded a larger area under the curve value compared with that from a previous study's scoring system. Therefore, in conclusion, since the calcium scoring system of the present study based on age and the maximum calcium arc obtained from OCT was specifically developed in the subjects with moderate and severe calcified lesions, it may be more accurate in predicting the risk of stent underexpansion in these patients.

Introduction

Coronary artery calcification is an integral process in atherogenesis, occurring in ≥90% of men and ≥67% of women >70 years of age (1). The pathogenesis of coronary calcification shares common pathways with bone formation, and eventually results in reduced vascular compliance, abnormal vasomotor responses and impaired myocardial perfusion. Calcified lesions are often harder to traverse and dilate, which pose higher risks of suboptimal stent deployment, angiographic complications and procedural failure (2). Coronary calcification presents a great challenge for percutaneous coronary intervention (PCI) and has an adverse impact on stent expansion and immediate treatment efficacy (3). The presence of calcified lesions strongly predicts the occurrence of stent thrombosis within a year of PCI and target lesion revascularization (4). Therefore, it is necessary to apply different approaches where calcified plaques are involved, prior to stent implantation, to achieve successful expansion.

Optical coherence tomography (OCT) utilizes near-infrared light directed at the vessel wall through a rotating single optical fiber coupled with an imaging lens within a short-monorail imaging sheath (5). By measuring the amplitude and time delay of the backscattered light, OCT generates high-resolution, cross-sectional and three-dimensional volumetric images of the coronary microstructure, which is an emerging intracoronary imaging modality that has been documented to accurately identify calcified lesions whilst also assessing the severity of calcification (6). A recent study reported that OCT-guided PCI

Correspondence to: Dr Jian Liu, Department of Cardiology, Peking University People's Hospital, 11 Xizhimen South Street, Xicheng, Beijing 100044, P.R. China
E-mail: drjianliu@163.com

*Contributed equally

Key words: calcified plaque, optical coherence tomography, stent underexpansion, percutaneous coronary intervention, calcium score

for calcified lesions resulted in improved stent expansion (7). Furthermore, Fujino *et al* (8) showed that maximum calcium angle, maximum calcium thickness and calcium length were independent predictors of stent underexpansion, and demonstrated that calcium lesions with a maximum angle of $>180^\circ$ (defined as 2 points), a maximum thickness of >0.5 mm (1 point) and a length >5 mm (1 point) may be at risk of stent underexpansion in patients with relatively mild calcification. It was also revealed that the lesions with calcium score of 0 to 3 had excellent stent expansion, whereas the lesions with a score of 4 had stent underexpansion. However, whether this previously established OCT-based scoring system can be applied to patients with moderate and severe calcified lesions remains unclear. Therefore, the present study aimed to develop a novel scoring system for the prediction of stent underexpansion in patients with moderate and severe calcified lesions.

Materials and methods

Study population. A total of 78 patients aged 18-90 years old (68.1% male) were screened for the present study. These patients were diagnosed with moderate or severe calcified coronary lesions using coronary angiography or OCT and underwent OCT-guided stent implantation at Peking University People's Hospital (Beijing, China) between January 2016 and July 2021. The degree of calcification on coronary angiography was classified according to the Mintz criteria (9). The lesion was considered to be moderate or severe calcified lesion on OCT if it had multiple complex calcium imaging features, such as a maximum calcium length >5 mm, a thickness >0.5 mm or an arc $>180^\circ$. The exclusion criteria were as follows: i) Lack of pre-procedure or post-stent OCT images; ii) in-stent restenosis and chronic total coronary artery occlusion; iii) incomplete OCT images, in which critical parameters could not be analyzed or quantified; and iv) poor image quality.

Study design. The present retrospective study aimed to develop a novel OCT-based calcium scoring system to predict stent underexpansion in moderate and severe calcified lesions. Medical records, including coronary angiography and OCT images, of the eligible patients were reviewed. The patient demographic information, clinical manifestation, past medical history, family history of coronary heart disease, laboratory examinations and PCI procedural characteristics were recorded. The present study was approved by the Ethics Committee of Peking University People's Hospital (Beijing, China; approval no. 2018PHB154-01) and was performed according to the principles of The Declaration of Helsinki.

PCI procedure and OCT imaging data acquisition. Coronary angiography and stent implantation were performed according to the standard protocols. Angiographic images were recorded using 5-6F angiographic imaging catheter or guiding catheter at 15 frames/sec by radiographic systems (Innova IGS 530, GE Medical Systems; Azurion7 M12, Philips Healthcare). Angiographic image runs at all standard projection views for each vessel were saved. The contrast medium was injected manually at a constant speed of approximately 4 ml/sec until

the distal vessel was filled. OCT imaging data were acquired using frequency-domain OCT (C7-XR™ or OPTIS™) and the Dragonfly™ Duo catheter (all purchased from Abbott Vascular; Abbott Pharmaceutical Co., Ltd.). After administration of intracoronary nitroglycerin, the Dragonfly™ Duo catheter was carefully advanced distal to the target lesion under fluoroscopic guidance. Then an automatic pullback OCT imaging was performed at a rate of 18 mm/sec (HD mode) or 36 mm/sec (S mode) throughout the entire lesion. During the session, contrast medium was flushed continuously with an injection rate of >5 ml/sec for the left coronary artery and >4 ml/sec for the right coronary artery depending on the vessel size. Based on the angiographic and OCT results, the decision of whether to perform rotational atherectomy (RA) or conventional angioplasty through cutting, scoring or using a non-compliant balloon prior to stent implantation was dependent on the discretion and skill of the surgeon. Briefly, in the presence of a multitude of complex calcium imaging features observable on the OCT image, an aggressive strategy of active RA was utilized followed by balloon angioplasty. Otherwise, the surgeon would first attempt a balloon angioplasty, followed by the RA procedure promptly in cases of inadequate balloon dilation. In cases of optimal balloon dilation or the apparent formation of a calcium crack (indicating adequate preparation of calcified lesions) after dilation on the OCT image, the stent would be implanted directly without the use of RA.

Quantitative coronary angiography (QCA) analysis. All angiography images were analyzed using the QAngio® XA software (version 7.3; Medis Medical Imaging Systems B.V.) by two independent interventional cardiologists (YM and QL), who were blinded to the clinical presentation and OCT results of the patients. The location, angulation and length of the calcified lesions, reference vessel diameter, minimum lumen diameter and diameter stenosis of the target lesions were assessed.

OCT imaging analysis. All OCT data were analyzed using Off-line Review Workstation software (version E.0.2; Abbott Vascular; Abbott Pharmaceutical Co., Ltd.) and based on procedures/guidelines described in dedicated expert consensus reports (10,11). This analysis was performed by two independent interventional cardiologists (YM and QL) who were blinded to the clinical and angiographic patient information.

In the present study, only lesions with a calcium arc $\geq 30^\circ$ were included. Lesions were considered to be two separate calcified lesions if there was >1 mm of non-calcified plaque between the two. If the boundary of the calcified lesions was not obvious, then the maximum visible thickness would be quantified. Superficial calcification would be defined if the distance between the lesion and the lumen was <100 μm . Stent edge dissection was defined as the interruption of surface continuity at the stent edge (within 5 mm distal and proximal to the stent). Stent malapposition would be defined if the longitudinal distance from the stent surface to the lumen was greater than the stent thickness, whereas tissue protrusion would be defined if there was protrusion of the tissue into the lumen following stent implantation. The percentage of stenosis area and diameter were defined as the minimum lumen area/mean reference lumen area and the mean lumen




Figure 1. Study flowchart. A total of 78 patients diagnosed with moderate or severe calcified coronary lesions using coronary angiography or OCT, who underwent OCT-guided PCI were screened for the present study. Of the 45 patients excluded, 25 were excluded due to lack of pre- or post-procedure OCT images, five showed in-stent restenosis, 10 had inadequate OCT images in which important parameters could not be quantified and five were unable to analyze due to poor image quality. In total, 33 patients (34 lesions) were finally included. OCT, optical coherence tomography; PCI, percutaneous coronary intervention.

diameter at the narrowest site/mean reference lumen diameter, respectively. The stent expansion percentage was calculated as the minimum stent area/mean reference vessel area $\times 100$. According to the determined stent expansion, patients were divided into the adequate (stent expansion $\geq 80\%$) and poor (stent expansion $< 80\%$) stent expansion groups (11,12).

Statistical analysis. Statistical analysis was performed using the SPSS software (version 24.0; IBM Corp.). Categorical variables are presented as n (%) and were compared using either the χ^2 or Fisher's exact tests, as appropriate. The Shapiro-Wilk test was performed to examine the normality of distribution. Continuous variables with a normal distribution are presented as the mean \pm standard deviation or are otherwise presented as the median and interquartile range. These variables were statistically compared using the unpaired Student's t-test or Mann-Whitney U test, as appropriate. Subsequently, the univariate logistic regression model was built and variables with $P < 0.10$ were included in the multivariate logistic regression model with a step-wise algorithm. The maximum calcium length factor, which was deemed to be clinically relevant, also entered into the multivariate analysis. Significant variables were then included into the final calcium scoring system.

Similar to the method of risk score establishment proposed in the Framingham Study, a novel calcium scoring system was developed by assigning weighted points for each variable (13). Receiver operating characteristic (ROC) curve analysis was performed to determine the optimal cut-off value for the novel scoring system for the prediction of stent underexpansion. The area under the curve (AUC) of the novel scoring system was compared with that proposed by Fujino *et al* (8) using the χ^2 test for two associated ROC curves. The inter-observer agreements for the OCT data were assessed by determining the intraclass correlation coefficients (ICC). $P < 0.05$ (two-sided) was considered to indicate a statistically significant difference.

Results

Clinical and procedural characteristics. After excluding 25 patients due to paucity of pre-procedure or post-stent OCT images, a total of 53 patients with moderate or severe calcified lesions, identified using coronary angiography or OCT, who underwent stent implantation with the guidance of OCT and had complete OCT images, were included into the present study. Among these patients, 33 (34 lesions) were

Table I. Clinical characteristics of the patients.

| Variables | Poor stent expansion (n=22) | Adequate stent expansion (n=11) | P-value |
|--|-----------------------------|---------------------------------|---------|
| Age, years | 70±10 | 59±8 | 0.003 |
| Male, n (%) | 12 (54.5) | 8 (72.7) | 0.456 |
| Body mass index, kg/m ² | 25.91±3.80 | 25.08±1.58 | 0.496 |
| Hypertension, n (%) | 16 (72.7) | 7 (63.6) | 0.696 |
| Diabetes, n (%) ^a | 10 (45.5) | 4 (36.4) | 0.719 |
| Hyperlipidaemia, n (%) | 16 (72.7) | 9 (81.8) | 0.687 |
| Chronic kidney disease, n (%) ^a | 2 (9.1) | 1 (9.1) | >0.999 |
| Smoking, n (%) | 8 (36.4) | 6 (54.5) | 0.534 |
| Family history of coronary heart disease, n (%) ^a | 4 (18.2) | 5 (45.5) | 0.121 |
| Prior percutaneous coronary intervention, n (%) ^a | 8 (36.4) | 4 (36.4) | >0.999 |
| Clinical diagnosis, n (%) ^a | | | >0.999 |
| Stable angina | 5 (22.7) | 2 (18.2) | |
| Unstable angina | 13 (59.1) | 7 (63.6) | |
| Acute myocardial infarction | 4 (18.2) | 2 (18.2) | |
| Left ventricular ejection fraction, % | 69.03±5.71 | 66.32±6.04 | 0.298 |
| Low density lipoprotein-cholesterol, mmol/l | 2.06±0.77 | 2.06±0.53 | 0.992 |
| Fasting plasma glucose, mmol/l | 5.08 (4.52-5.58) | 5.07 (4.72-6.10) | 0.620 |
| Estimated glomerular filtration rate, ml/min/1.73 m ² | 82.16 (71.37-93.25) | 93.95 (90.52-104.00) | 0.036 |

^aCompared using Fisher's exact test. Data are presented as the number (%), mean ± standard deviation or median (interquartile range).

finally included according to the aforementioned exclusion criteria (Fig. 1).

The mean age of the patients was 67±10 years and 20 of the patients were male. Furthermore, 23 (69.7%) patients had coexisting hypertension and 25 (75.8%) had hyperlipidemia. Of all 33 patients, 7 (21.2%) were diagnosed with stable angina pectoris, whereas 26 (78.8%) had acute coronary syndrome. Poor stent expansion occurred in 22 patients (23 lesions). The patients in the poor stent expansion group were significantly older compared with those in the adequate stent expansion group (70±10 vs. 59±8 years, respectively; P=0.003). In addition, the estimated glomerular filtration rate was significantly lower in the poor stent expansion group compared with that in the adequate stent expansion group (82.16 vs. 93.95; P=0.036; Table I). No significant difference was observed between the two groups with regards to the remaining clinical characteristics (Table I).

RA was performed in the modification of 19 (55.9%) lesions. Furthermore, 13 (38.2%) lesions were treated with scoring whereas 13 (38.2%) were treated with a non-compliant balloon prior to stent deployment. The median number of stents implanted was two and the median total length of stents was 44 mm. Compared with that in patients in the adequate stent expansion group [2 (18.2%)], the rate of RA performed during PCI was significantly higher in the patients in the poor stent expansion group [17 (73.9%); P=0.003; Table II]. There was no significant difference regarding the usage of balloons and stents between the two stent expansion groups (Table II).

Imaging analysis of the calcified plaques. The majority of patients had multi-vessel disease. The target lesions were mainly located in the left anterior descending artery. The prevalence of moderate and severe coronary calcification as assessed by angiography was up to 88.2% (30 lesions). All parameters of QCA analysis, including target vessel, degree of calcification, angulation of lesions, calcium length, minimum lumen diameter, minimum stent diameter and reference vessel diameter were comparable between the two stent expansion groups (Table III).

There were high levels of similarity between the two observers for the interpretation of the OCT images and the assessment of the maximum calcium arc (ICC=0.877), thickness (ICC=0.874) and length (ICC=0.968) (data not shown). The lesions all manifested as superficial calcifications, with a median maximum calcium arc of 230°, median maximum calcium length of 25.10 mm and an average maximum calcium thickness of 1.18 mm. The overall final post-PCI median stent expansion was 70.74%. Furthermore, the maximum calcium arc (299 vs. 142°; P=0.001; Table IV) and thickness (1.24 vs. 1.04 mm; P=0.029; Table IV) were significantly larger in the poor stent expansion group compared with those in the adequate expansion group, whereas the pre-stent diameter stenosis determined using OCT was significantly lower (52.76 vs. 66.29%; P=0.038; Table IV). The proportion of stent malapposition was significantly higher in the poor stent expansion group compared with that in the adequate stent expansion group (60.9 vs. 9.1%; P=0.011; Table IV), whereas the mean stent expansion

Table II. Procedural characteristics of the lesions.

| Variables | Poor stent expansion (n=23) ^a | Adequate stent expansion (n=11) ^a | P-value |
|---|--|--|---------|
| Scoring balloon, n (%) ^b | 11 (47.8) | 2 (18.2) | 0.245 |
| Maximum pressure, atm | 14.00 (11.00-15.00) | 12.00 (12.00-12.00) | 0.545 |
| Maximum diameter, mm | 2.75 (2.50-2.88) | 2.50 (2.38-2.62) | 0.351 |
| Non-compliant balloon, n (%) ^b | 11 (47.8) | 2 (18.2) | 0.245 |
| Maximum pressure, atm | 18.00 (17.00-24.00) | 13.50 (12.75-14.25) | 0.091 |
| Maximum diameter, mm | 2.50 (2.50-2.50) | 2.25 (2.12-2.38) | 0.098 |
| Semi-compliant balloon, n (%) | 18 (78.3) | 8 (72.7) | >0.999 |
| Maximum pressure, atm | 16.00 (14.00-16.00) | 15.00 (13.50-16.50) | 0.686 |
| Maximum diameter, mm | 2.50 (2.12-2.50) | 2.50 (2.38-2.50) | 0.885 |
| RA, n (%) ^b | 17 (73.9) | 2 (18.2) | 0.003 |
| No. of RA procedures ^b | 5 (3-7) | 5 (4-5) | 0.893 |
| Maximum burr size, mm | 1.50 (1.38-1.50) | 1.50 (1.50-1.50) | 0.725 |
| Speed of burr, x10 ⁴ r/min | 15.00 (15.00-16.00) | 16.60 (15.90-17.30) | 0.212 |
| Stent | | | |
| Number of stents ^b | 2 (1-3) | 1 (1-2) | 0.126 |
| Total stent length, mm | 48.00 (37.50-69.00) | 32.00 (28.00-57.00) | 0.197 |
| Maximum diameter, mm | 3.00 (2.75-3.00) | 2.75 (2.50-3.00) | 0.165 |

^aRepresents the number of lesions. ^bCompared using Fisher's exact test. Data are presented as the number (%) or median (interquartile range). RA, rotational atherectomy.

was significantly lower in the poor stent expansion group (63.39±12.72 vs. 86.10±4.59%; P<0.001; Table IV).

Development of the novel calcium scoring system and comparisons with the previous scoring system. In the univariate analysis, age, maximum calcium arc and diameter stenosis were found to be significant predictors of stent underexpansion (P<0.05). Age [odds ratio (OR), 1.173; 95% CI, 1.036-1.438; P=0.042] and maximum calcium arc (OR, 1.023; 95% CI, 1.008-1.050; P=0.021) were demonstrated to be independent predictors of stent underexpansion in the multivariate logistic regression model (Table V). A novel calcium scoring system was established as follows: (0.16 x age) + (0.03 x maximum calcium arc), with the points of each variable assigned based on β-coefficients in the aforementioned multivariate analysis. In order to simplify the model and facilitate the calculation, each β-coefficient was rounded from 0.159 to 0.16, and from 0.023 to 0.03.

Subsequently, the optimal threshold for the prediction of stent underexpansion was identified using ROC curve analysis. The optimal cut-off value for the scoring system was determined to be 16.87 (sensitivity, 0.870; specificity, 0.909; AUC, 0.925; 95% CI, 0.836-1.014; P<0.001; Fig. 2). Representative OCT images are presented in Fig. 3. The pre-procedure OCT image of a 67-year-old patient showed severe coronary calcification with a maximum calcium arc of 290° at the narrowest site. The calculated calcium score of the lesion was 19.42 according to the novel calcium scoring system, which predicted the occurrence of stent underexpansion. The final post-stent OCT image indicated poor stent expansion as anticipated with a minimum stent area of 4.38 mm².




Figure 2. Receiver operating characteristic curve analysis of the novel calcium scoring system and the previously widely used system devised by Fujino *et al* (8). The optimal threshold of the novel calcium scoring system for predicting stent underexpansion was 16.87 with a sensitivity of 0.870, a specificity of 0.909 and an AUC of 0.925 (95% CI, 0.836-1.014; P<0.001). The AUC for the novel calcium scoring system was significantly larger (0.925 vs. 0.706; P=0.002). AUC, area under the curve.

A previous calcium scoring system developed by Fujino *et al* (8) was also applied, whereby the corresponding score for each patient was calculated before ROC curve analysis was performed to investigate the efficacy for the prediction of poor stent expansion. Compared with that of the Fujino *et al* (8) scoring system, the novel system in the present study, which incorporated age and the maximum calcium arc, was found to be superior for the prediction of stent underexpansion in the present study population, with a significantly larger AUC for the ROC analysis (0.925 vs. 0.706, respectively; P=0.002; Fig 2).

Table III. Quantitative coronary angiography analyses of calcified lesions.

| A, Pre-procedure | | | |
|---|--|--|---------|
| Variables | Poor stent expansion (n=23) ^a | Adequate stent expansion (n=11) ^a | P-value |
| Multivessel disease, n (%) | 19 (82.6) | 11 (100.0) | 0.280 |
| Target vessel, n (%) ^b | | | >0.999 |
| Left anterior descending | 19 (82.6) | 9 (81.8) | |
| Left circumflex | 1 (4.3) | 0 (0.0) | |
| Right coronary artery | 3 (13.0) | 2 (18.2) | |
| Degree of calcification, n (%) ^b | | | 0.362 |
| None or mild | 2 (8.7) | 2 (18.2) | |
| Moderate | 14 (60.9) | 8 (72.7) | |
| Severe | 7 (30.4) | 1 (9.1) | |
| Bifurcation, n (%) ^b | 2 (8.7) | 4 (36.4) | 0.070 |
| Angulation, n (%) ^b | | | 0.203 |
| ≤90° | 22 (95.7) | 10 (90.9) | |
| >90° | 1 (4.3) | 1 (9.1) | |
| Calcium length, mm | 35.77±20.66 | 26.03±12.34 | 0.227 |
| RVD, mm | 2.63 (2.38-2.83) | 2.54 (2.32-2.64) | 0.597 |
| Minimum lumen diameter, mm | 1.24±0.42 | 1.33±0.33 | 0.579 |
| Diameter stenosis, % | 53.43±13.27 | 48.02±12.92 | 0.290 |
| B, Post-procedure | | | |
| Variables | Poor stent expansion (n=23) ^a | Adequate stent expansion (n=11) ^a | P-value |
| RVD, mm | 2.39 (2.12-2.53) | 2.50 (2.26-2.94) | 0.155 |
| Minimum stent diameter, mm | 2.19 (2.02-2.44) | 2.33 (2.02-2.74) | 0.382 |
| Diameter stenosis, % | 11.58 (6.45-15.02) | 10.71 (6.80-15.51) | 0.887 |

^aRepresents the number of lesions. ^bCompared using Fisher's exact test. Data are presented as the number (%), mean ± standard deviation or median (interquartile range). RVD, reference vessel diameter.

Discussion

The present study established a novel scoring system to effectively predict stent underexpansion in moderate and severe calcified lesions. The procedural characteristics and intracoronary imaging data of patients with moderate and severe coronary calcification who had undergone PCI were retrospectively analyzed. The main findings of the present study were as follows: i) The maximum calcium arc of lesions and patient age were independent predictors of stent underexpansion in patients with moderate and severe coronary calcification; and ii) the calcium scoring system based on these parameters may accurately predict the risk of stent underexpansion and guide the strategy of lesion modification, such as RA, in patients with moderate and severe coronary calcification.

It has previously been reported that the occurrence of stent underexpansion is increased in patients with severe calcified lesions, where incomplete stent expansion is known to be a common risk factor for stent thrombosis and in-stent

restenosis (14-16). Severe coronary calcification is also an independent predictor of poor prognosis and increases the mortality rate (10.8 vs. 4.4%; $P < 0.001$) or 1-year other adverse cardiac events defined as cardiac death, myocardial infarction, and target vessel revascularization after the treatment procedure (24.4 vs. 4.7%; $P < 0.001$) (17,18). Therefore, a more aggressive strategy of lesion modification is needed prior to stent deployment to achieve efficient interventional treatment for this condition (19). Matsuhiro *et al* (20) previously reported that maximum calcium thickness $< 880 \mu\text{m}$ was a useful predictor of acceptable stent expansion (defined as 80% expansion) in moderate calcified lesions. Furthermore, Maejima *et al* (21) demonstrated that larger calcium arcs and a lower calcium thickness were associated with the formation of calcium cracks, which are important determinants of optimal stent expansion. However, in the present study it was demonstrated that the maximum calcium arc and the age of patients, but not the thickness of the lesions, actually had the more significant impact on stent expansion.

Table IV. Optical coherence tomography data of calcified lesions.

| A, Pre-stent | | | |
|---|--|--|---------|
| Variables | Poor stent expansion (n=23) ^a | Adequate stent expansion (n=11) ^a | P-value |
| Superficial calcium, n (%) | 23 (100) | 11 (100) | - |
| Maximum calcium length, mm | 33.15 (15.62-40.20) | 20.65 (14.78-25.10) | 0.094 |
| Maximum calcium arc, degree | 299 (205-345) | 142 (104-216) | 0.001 |
| Maximum calcium thickness, mm | 1.24±0.23 | 1.04±0.25 | 0.029 |
| Minimum lumen area, mm ² | 1.64 (1.16-2.28) | 2.26 (1.65-2.71) | 0.217 |
| Reference vessel area, mm ² | 6.47 (5.08-8.20) | 5.30 (5.11-6.92) | 0.429 |
| Diameter stenosis, % | 52.76 (45.59-58.58) | 66.29 (55.66-71.74) | 0.038 |
| Area stenosis, % | 81.73 (80.06-83.37) | 84.11 (82.44-86.80) | 0.063 |
| B, Post-stent | | | |
| Variables | Poor stent expansion (n=23) ^a | Adequate stent expansion (n=11) ^a | P-value |
| Tissue prolapse, n (%) ^b | 1 (4.3) | 0 (0) | >0.999 |
| Stent edge dissection, n (%) | 0 (0) | 0 (0) | - |
| Stent malapposition, n (%) ^b | 13 (60.9) | 1 (9.1) | 0.011 |
| Minimum stent area, mm ² | 4.25 (3.96-5.62) | 5.42 (4.60-6.77) | 0.146 |
| Reference vessel area, mm ² | 8.02 (6.43-9.09) | 5.90 (5.33-8.64) | 0.217 |
| Stent expansion, % | 63.39±12.72 | 86.10±4.59 | <0.001 |

^aRepresents the number of lesions. ^bCompared using Fisher's exact test. Data are presented as the number (%), mean ± standard deviation or median (interquartile range).

Table V. Univariate and multivariate logistic regression model of stent underexpansion.

| A, Univariate analysis | | | | |
|-----------------------------|---------------|------------|-------------|---------|
| Variables | β-coefficient | Odds ratio | 95% CI | P-value |
| Age, years | 0.133 | 1.143 | 1.044-1.297 | 0.013 |
| Maximum calcium arc, degree | 0.019 | 1.019 | 1.008-1.035 | 0.004 |
| Diameter stenosis, % | -0.063 | 0.939 | 0.872-0.988 | 0.048 |
| B, Multivariate analysis | | | | |
| Variables | β-coefficient | Odds ratio | 95% CI | P-value |
| Age, years | 0.159 | 1.173 | 1.036-1.438 | 0.042 |
| Maximum calcium arc, degree | 0.023 | 1.023 | 1.008-1.050 | 0.021 |

The inconsistency in the inclusion criteria among these studies may partially explain the different results observed. Furthermore, the present study proposed a novel calcium scoring system based on the aforementioned parameters, which displayed high accuracy in predicting the risk of stent underexpansion.

OCT confers superior improved capability compared with coronary angiography for the detection of calcium, with a

sensitivity ranging between 95 and 96% and a specificity of 97% (22,23). In addition, compared with visual angiographic assessment alone, intracoronary OCT images can provide additional information on the parameters associated with the calcification severity of target lesions, such as maximum calcium arc, thickness, depth and longitudinal length (6). In the present study, the maximum calcium arc determined from the




Figure 3. Representative OCT images. (A) Lesion manifested as severe calcified stenosis, which had a calcium arc of 290° (arrowheads) and a minimum lumen area of 1.53 mm² as determined on the pre-stent OCT image. The calcium score of the lesion was 19.42. (B) Following stent implantation, the final OCT image indicated poor stent expansion with a minimum stent area of 4.38 mm². OCT, optical coherence tomography.

OCT images was demonstrated to be a potential independent predictor of stent underexpansion, whereas parameters from QCA were not. These results highlighted the importance and value of intravascular imaging modality for severe calcified lesions in the interventional strategy-making process. The Society for Cardiovascular Angiography and Interventions position statement published in 2020 recommended that moderate and severe calcium observed on coronary angiography, as well as inadequate balloon expansion during lesion preparation before stent implantation, should be evaluated by intravascular imaging (24). In the present study, angiographically visible moderate and severe calcified lesions were found in 30 (88.2%) lesions. By contrast, the remaining four lesions, which showed mild calcification on angiography, had poor balloon expansion during lesion preparation. Therefore, OCT evaluation of all lesions was performed according to the recommendation in the aforementioned 2020 position statement.

The results of a previous study reported that for lesions with none/mild calcification the rate of major adverse cardiac events at 1-year is only 8.3%, whereas this increases to 14.6 and 17.7% for moderate and severe calcified lesions, respectively (25). Fujino *et al* (8) previously developed an OCT-based calcium scoring system to identify lesions, which may be at risk of stent underexpansion and benefit from plaque modification prior to stent implantation. In this previous study, only 29.7% of the patients enrolled had moderate and/or severe calcified lesions as assessed using angiography. Furthermore, the OCT characteristics of the patients tended to be relatively mild calcification. Patients treated with RA or scoring balloon were excluded. By contrast, in the present study 88.2% of the lesions had angiographically visible moderate and severe calcification. These two scoring systems were also compared. The present study's population with moderate and severe coronary calcification demonstrated that this newly-established system exhibited

an improved performance compared with the widely-used Fujino *et al* (8) system in predicting the immediate therapeutic outcome of PCI. The comparison of predictive performance between the novel and the Fujino *et al* (8) system in a patient population with mild calcified plaques may be the aim of future studies.

It is noteworthy that the proportion of RA during PCI was significantly higher in patients with poor stent expansion, which may potentially be the result of lesion calcification being more severe in this group of patients. Furthermore, there was a discrepancy between groups with adequate or poor stent expansion in the present study. Compared with the patients who had adequate stent expansion, patients in the poor stent expansion group were significantly older. The maximum calcium arc and thickness were also larger in the poor stent expansion group. However, the variables showing significant difference between the two groups and potential confounding factors, such as maximum calcium length, were included into the univariate and multivariate logistic regression analyses, to limit the influence of this discrepancy and to ensure the accuracy of the results.

The present study also had several limitations. The present study was retrospective, where leaving the interventional strategy to the surgeons' discretion may have affected the stent expansion and the final analysis. In addition, the relatively small number of patients were enrolled, which made the conclusion drawn from the study weaker. There was also a lack of a specific cohort to validate the accuracy of the present calcium scoring system to predict stent underexpansion. Furthermore, as the OCT images were not read and analyzed by the same interventional cardiologist at different times, the intra-observer concordance for the assessment of the OCT data could not be assessed. Information regarding peri-procedural complications, such as coronary dissection or perforation and the balloon used for post-dilatation, was not available. Finally, the capability of

the calcium scoring system to predict the long-term clinical outcome post-PCI in patients with moderate and severe calcified lesions remains to be elucidated.

In conclusion, the novel OCT-based calcium scoring system in the present study demonstrated a high accuracy for the prediction of the occurrence of stent underexpansion in patients with moderate and severe coronary calcification. However, the present system requires further validation in a larger cohort.

Acknowledgements

The authors would like to thank Professor Zhuang Tao (Shanghai MedStat Clinical Research Institute, Shanghai, China) for his help in the statistical analysis of the present study.

Funding

The present study was supported by the National Natural Science Foundation of China (grant no. 11832003 and 81970294) and Peking University People's Hospital Scientific Research Development Funds (grant no. RDL-2020-11).

Availability of data and materials

The datasets used and/or analyzed during the current study are available from the corresponding author on reasonable request.

Authors' contributions

CH and LY collected, analyzed and interpreted the patient data. ZX and HL contributed to collection and analysis of the data. YM and QL performed the analysis and interpretation of angiographic and OCT data. ML and HZ confirm the authenticity of all the raw data and revised the manuscript critically for important intellectual content. CL and JL designed the present study. YM, QL, ML, HZ and JL performed the operation. All authors read and approved the final manuscript.

Ethics approval and consent to participate

The present study was approved by the Ethics Committee of Peking University People's Hospital (Beijing, China; approval no. 2018PHB154-01) and conducted according to the principles of the Declaration of Helsinki. Since this clinical study was a retrospective analysis of the information of previous cases, without direct contact with the subjects and subject privacy protection, the risk borne by the subjects was not greater than the minimum risk. The Ethics Committee of Peking University People's Hospital waived the requirement for informed consent.

Patient consent for publication

Not applicable.

Competing interests

The authors declare that they have no competing interests.

References

1. Madhavan MV, Tarigopula M, Mintz GS, Maehara A, Stone GW and Généreux P: Coronary artery calcification: Pathogenesis and prognostic implications. *J Am Coll Cardiol* 63: 1703-1714, 2014.
2. Tomey MI and Sharma SK: Interventional options for coronary artery calcification. *Curr Cardiol Rep* 18: 12, 2016.
3. Kobayashi Y, Okura H, Kume T, Yamada R, Kobayashi Y, Fukuhara K, Koyama T, Nezu S, Neishi Y, Hayashida A, *et al*: Impact of target lesion coronary calcification on stent expansion. *Circ J* 78: 2209-2214, 2014.
4. Généreux P, Madhavan MV, Mintz GS, Maehara A, Palmerini T, Lasalle L, Xu K, McAndrew T, Kirtane A, Lansky AJ, *et al*: Ischemic outcomes after coronary intervention of calcified vessels in acute coronary syndromes. Pooled analysis from the HORIZONS-AMI (harmonizing outcomes with revascularization and stents in acute myocardial infarction) and ACUITY (acute catheterization and urgent intervention triage strategy) TRIALS. *J Am Coll Cardiol* 63: 1845-1854, 2014.
5. Ali ZA, Karimi Galougahi K, Mintz GS, Maehara A, Shlofmitz RA and Mattesini A: Intracoronary optical coherence tomography: State of the art and future directions. *EuroIntervention* 17: e105-e123, 2021.
6. Mehanna E, Bezerra HG, Prabhu D, Brandt E, Chamié D, Yamamoto H, Attizzani GF, Tahara S, Van Ditzhuijzen N, Fujino Y, *et al*: Volumetric characterization of human coronary calcification by frequency-domain optical coherence tomography. *Circ J* 77: 2334-2340, 2013.
7. Kobayashi N, Ito Y, Yamawaki M, Araki M, Obokata M, Sakamoto Y, Mori S, Tsutsumi M, Honda Y, Makino K, *et al*: Optical coherence tomography-guided versus intravascular ultrasound-guided rotational atherectomy in patients with calcified coronary lesions. *EuroIntervention* 16: e313-e321, 2020.
8. Fujino A, Mintz GS, Matsumura M, Lee T, Kim SY, Hoshino M, Usui E, Yonetsu T, Haag ES, Shlofmitz RA, *et al*: A new optical coherence tomography-based calcium scoring system to predict stent underexpansion. *EuroIntervention* 13: e2182-e2189, 2018.
9. Mintz GS, Popma JJ, Pichard AD, Kent KM, Satler LF, Chuang YC, Ditrano CJ and Leon MB: Patterns of calcification in coronary artery disease. A statistical analysis of intravascular ultrasound and coronary angiography in 1155 lesions. *Circulation* 91: 1959-1965, 1995.
10. Prati F, Guagliumi G, Mintz GS, Costa M, Regar E, Akasaka T, Barlis P, Tearney GJ, Jang IK, Arbustini E, *et al*: Expert review document part 2: Methodology, terminology and clinical applications of optical coherence tomography for the assessment of interventional procedures. *Eur Heart J* 33: 2513-2520, 2012.
11. Tearney GJ, Regar E, Akasaka T, Adriaenssens T, Barlis P, Bezerra HG, Bouma B, Bruining N, Cho JM, Chowdhary S, *et al*: Consensus standards for acquisition, measurement, and reporting of intravascular optical coherence tomography studies: A report from the International working group for intravascular optical coherence tomography standardization and validation. *J Am Coll Cardiol* 59: 1058-1072, 2012.
12. Räber L, Mintz GS, Koskinas KC, Johnson TW, Holm NR, Onuma Y, Radu MD, Joner M, Yu B, Jia H, *et al*: Clinical use of intracoronary imaging. Part 1: Guidance and optimization of coronary interventions. An expert consensus document of the European association of percutaneous cardiovascular interventions. *Eur Heart J* 39: 3281-3300, 2018.
13. Sullivan LM, Massaro JM and D'Agostino RB Sr: Presentation of multivariate data for clinical use: The Framingham study risk score functions. *Stat Med* 23: 1631-1660, 2004.
14. Mosseri M, Satler LF, Pichard AD and Waksman R: Impact of vessel calcification on outcomes after coronary stenting. *Cardiovasc Revasc Med* 6: 147-153, 2005.
15. Fujii K, Carlier SG, Mintz GS, Yang YM, Moussa I, Weisz G, Dangas G, Mehran R, Lansky AJ, Kreps EM, *et al*: Stent underexpansion and residual reference segment stenosis are related to stent thrombosis after sirolimus-eluting stent implantation: An intravascular ultrasound study. *J Am Coll Cardiol* 45: 995-998, 2005.
16. Kastrati A, Dibra A, Mehilli J, Mayer S, Pinićek S, Pache J, Dirschinger J and Schömig A: Predictive factors of restenosis after coronary implantation of sirolimus- or paclitaxel-eluting stents. *Circulation* 113: 2293-2300, 2006.
17. Bourantas CV, Zhang YJ, Garg S, Iqbal J, Valgimigli M, Windecker S, Mohr FW, Silber S, Vries TD, Onuma Y, *et al*: Prognostic implications of coronary calcification in patients with obstructive coronary artery disease treated by percutaneous coronary intervention: A patient-level pooled analysis of 7 contemporary stent trials. *Heart* 100: 1158-1164, 2014.

18. Sharma SK, Bolduan RW, Patel MR, Martinsen BJ, Azemi T, Giugliano G, Resar JR, Mehran R, Cohen DJ, Popma JJ and Waksman R: Impact of calcification on percutaneous coronary intervention: MACE-Trial 1-year results. *Catheter Cardiovasc Interv* 94: 187-194, 2019.
19. Tang Z, Bai J, Su SP, Lee PW, Peng L, Zhang T, Sun T, Nong JG, Li TD and Wang Y: Aggressive plaque modification with rotational atherectomy and cutting balloon for optimal stent expansion in calcified lesions. *J Geriatr Cardiol* 13: 984-991, 2016.
20. Matsuhiro Y, Nakamura D, Shutta R, Yanagawa K, Nakamura H, Okamoto N, Egami Y, Sakata Y, Nishino M and Tanouchi J: Maximum calcium thickness is a useful predictor for acceptable stent expansion in moderate calcified lesions. *Int J Cardiovasc Imaging* 36: 1609-1615, 2020.
21. Maejima N, Hibi K, Saka K, Akiyama E, Konishi M, Endo M, Iwahashi N, Tsukahara K, Kosuge M, Ebina T, *et al*: Relationship between thickness of calcium on optical coherence tomography and crack formation after balloon dilatation in calcified plaque requiring rotational atherectomy. *Circ J* 80: 1413-1419, 2016.
22. Yabushita H, Bouma BE, Houser SL, Aretz HT, Jang IK, Schlendorf KH, Kauffman CR, Shishkov M, Kang DH, Halpern EF and Tearney GJ: Characterization of human atherosclerosis by optical coherence tomography. *Circulation* 106: 1640-1645, 2002.
23. Wang X, Matsumura M, Mintz GS, Lee T, Zhang W, Cao Y, Fujino A, Lin Y, Usui E, Kanaji Y, *et al*: In vivo calcium detection by comparing optical coherence tomography, intravascular ultrasound, and angiography. *JACC Cardiovasc Imaging* 10: 869-879, 2017.
24. Riley RF, Henry TD, Mahmud E, Kirtane AJ, Brilakis ES, Goyal A, Grines CL, Lombardi WL, Maran A, Rab T, *et al*: SCAI position statement on optimal percutaneous coronary interventional therapy for complex coronary artery disease. *Catheter Cardiovasc Interv* 96: 346-362, 2020.
25. Copeland-Halperin RS, Baber U, Aquino M, Rajamanickam A, Roy S, Hasan C, Barman N, Kovacic JC, Moreno P, Krishnan P, *et al*: Prevalence, correlates, and impact of coronary calcification on adverse events following PCI with newer-generation DES: Findings from a large multiethnic registry. *Catheter Cardiovasc Interv* 91: 859-866, 2018.



This work is licensed under a Creative Commons Attribution-NonCommercial-NoDerivatives 4.0 International (CC BY-NC-ND 4.0) License.